

Contacts:

Stéphanie SAMMARTANO
Certification Engineer
stephanie.sammartano@afnor.org
Phone : +33 (0)1 41 62 62 39

Audrey VERNEL
Operations supervisor
audrey.vernel@afnor.org
Phone : +33 (0)1 41 62 60 63

Ref.: SSO/AYV/NF102/Clients/OXOID Thermo Fisher Scientific/
Avis BT_SureTect Salmonella_2013-11-04_(V)

Subject: NF VALIDATION mark

Oxoid Ltd
Mr. Jonathan CLOKE
Wade Road
RG24 8PW Basingstoke
Hampshire
UK

La Plaine Saint-Denis, November 5th, 2013

Dear Sir,

Following the positive agreement expressed the November 4th, 2013, by the Technical Board "Food microbiology" of the NF VALIDATION mark (NF102), I beg to inform you that the **NF VALIDATION certification is granted** to OXOID Thermo Fisher Scientific company for the following alternative analysis method:

Thermo Scientific SureTect *Salmonella* species PCR Assay
Validated for the detection of *Salmonella* in non-seasoned raw ground beef meat
by comparison to EN ISO 6579

Certificate reference n# UNI 03/07-11/13, with end of validity 4-Nov.-2017

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.

Managing Director
Florence MÉAUX



Contacts:

Stéphanie SAMMARTANO
Ingénieure certification
stephanie.sammartano@afnor.org
Tél : +33 (0)1 41 62 62 39

Audrey VERNEL
Chargée de clientèle
audrey.vernel@afnor.org
Tél : +33 (0)1 41 62 60 63

Réf.: SSO/AYV/NF102/Clients/OXOID Thermo Fisher Scientific/
Avis BT_SureTect Salmonella_2013-11-04_(V)

Objet : Marque NF VALIDATION

Oxoid Ltd
Monsieur Jonathan CLOKE
Wade Road
RG24 8PW Basingstoke
Hampshire
UK

La Plaine Saint-Denis, le 5 novembre 2013

Monsieur,

Comme suite à l'avis positif exprimé le 4 novembre 2013 par le Bureau Technique « Microbiologie agroalimentaire » de la marque NF VALIDATION (NF102), j'ai l'honneur de vous annoncer que le **droit d'usage de la marque NF VALIDATION est accordée** à la société OXOID Thermo Fisher Scientific pour la méthode alternative d'analyse suivante :

Thermo Scientific SureTect *Salmonella* species PCR Assay

Validée pour la détection des *Salmonella*
dans les viandes crues de bœuf hachées (non assaisonnées)
par comparaison à la norme NF EN ISO 6579

Certifiée sous le n° d'attestation UNI 03/07-11/13, avec pour fin de validité le 04-nov.-2017

Un courrier complet de conclusions mentionnant d'éventuelles réserves prononcées par le Bureau Technique, vous sera adressé prochainement. Dans ce cas, celles-ci devront être prises en compte sans délai.

Je vous prie d'agréer, Monsieur, mes salutations distinguées.

Directrice Générale
Florence MÉAUX