









Acid Phosphatase Reagent

ADDITION OF COMPONENT 1		Pipette 1.5 ml of component 1 into the vial with the colour reagent component 2.	1 min	
VORTEX		Vortex until a homogeneous mixture is achieved.	1 min	
COOL		Protecting it from light, cool the vial for 1 hour at 2...8 °C.	60 min	
CENTRIFUGE		Then centrifuge the vial for 10 minutes at 6,000 RPM.	10 min	
TRANSFER THE SUPERNATANT		Transfer 750 µl of the supernatant into a fresh vial.	2 min	
ADDITION OF ACID PHOSPHATASE REAGENT REACTION ON FILTER PAPER		To detect acid phosphatase, put 2-3 drops of the reagent onto filter paper and rub a suspect (<i>Clostridium perfringens</i>) colony into it. If there is a positive reaction, the colony changes colour after 3-4 minutes and ends up dark red to brown red.	5 min	